



CERTIFICATE OF CALIBRATION

Issued by: **AZMOON SANJESH ABZAR**

Date of issue: 1393-08-28

Certificate NO: **93-C-24061**

NO32-POONAK BLVD-ASHRAFI ESFAHANI ST-TEHRAN-IRAN

Telefax: +9821 44445009 / +9821 44401554

Email: asacalibration@yahoo.com

<http://www.asa-calibration.com>

شرکت مهندسی آسا
آزمون سنجش ابزار

page 1 of 2 pages
Calibration date: 1393-08-27
Next calibration: 1394-08-27

APPLICANT: پشیران صنعت ویرا

INSTRUMENT DESCRIPTION: Instrument: Temperature-Humidity meter
Serial NO: ---
Model: TM-1250

Manufacturer: Tika
IDentification NO: ---

BASIS OF CALIBRATION: Based on internal procedure LTW-12-00

CALIBRATION METHOD: Comparison method- By putting the instrument in the climatic chamber and comparison between the reference values and the under calibration.
The temperature scale use in this laboratory is based on international temperature scale 1990.

ENVIRONMENT: The temperature of the working volume should be maintained within $23 \pm 2^{\circ}\text{C}$. Variations in temperature at any position should not exceed 2°C per day and 1°C per hour and relative humidity must be between 30% ~ 60%.

REFERENCE EQUIPMENTS: Reference temperature-Humidity meter/Serial No:03026488/Certificate NO:S2939/19
Climatic chamber with air circulation system

TRACEABILITY: Traceable to *NIST* by certificate No.S2939/19
(*NIST*: National Institute Standard and Technology)

QUALITY SYSTEM: Based on ISO / IEC 17025:2005

LABORATORY STAMP

CALIBRATION TECHNICIAN

APPROVED SIGNATORY



B.Heidari

A.Moosavi

ID NO. LQS-F363-00